



Bids and Awards Committee

BID BULLETIN NO. 2

PROJECT : Procurement of Technical and Scientific Equipment
LOCATION : NBI MAIN OFFICE, TAFT AVENUE, MANILA
DATE : 27 December 2011

Notice is hereby given to the following registered bidders in connection with the Resolutions of the NBI Bids and Awards Committee dated 21 December 2011, pertinent sections of which are quoted herein, relative to the procurement of **Technical and Scientific Equipment: Lot 5 – REFRIGERATED CENTRIFUGE AND REAL TIME PCR:**

1. For MACARE MEDICALS, INC.:

WHEREAS, in compliance with said directive, the TWG conducted the required post-qualification, during which the bid documents of MACARE MEDICALS, INC. were found to be non-compliant with the Technical Qualifications/Specifications of the bid pursuant to the provision of RA 9184:

- No Certificate of Validation that the Machine is used for Human Identification (hid) in accordance with the guidelines of the DNA Advisory Board, SWGDAM, or other similar bodies;
- The proposal does not include start-up consumables, HID installation and quantitation kits for internal validation;
- The refrigerated centrifuge offered is non-compliant with the technical specifications;

NOW, THEREFORE, the Bids and Awards Committee (BAC), through the Undersigned Members RESOLVED, AS IT HEREBY RESOLVES, THAT MACARE MEDICALS, INC. FAILED TO MEET THE TECHNICAL REQUIREMENTS OF THE BID.

2. For PHILAB INDUSTRIES INC.:

WHEREAS, in compliance with said directive, the TWG conducted the required post-qualification, during which the bid documents of PHILAB INDUSTRIES INC. were found to be non-compliant with the Technical Qualifications/Specifications of the bid pursuant to the provision of RA 9184:

- No Certificate of Validation that the Machine is used for Human Identification (hid) in accordance with the

guidelines of the DNA Advisory Board, SWGDAM, or other similar bodies;

- The refrigerated centrifuge offered is non-compliant with the technical specifications;

NOW, THEREFORE, the Bids and Awards Committee (BAC), through the Undersigned Members RESOLVED, AS IT HEREBY RESOLVES, THAT PHILAB INDUSTRIES INC. FAILED TO MEET THE TECHNICAL REQUIREMENTS OF THE BID.

3. FIL-ANASERVE, INC.:

WHEREAS, in compliance with said directive, the TWG conducted the required post-qualification, during which the bid documents of FIL-ANASERVE, INC. were found to be non-compliant with the Technical Qualifications/Specifications of the bid pursuant to the provision of RA 9184:

- The machine offered Real Time PCR 7500 Fast is not yet validated for Human Identification, hence no certification can be issued to this effect;

NOW, THEREFORE, the Bids and Awards Committee (BAC), through the Undersigned Members RESOLVED, AS IT HEREBY RESOLVES, THAT FIL-ANASERVE, INC. FAILED TO MEET THE TECHNICAL REQUIREMENTS OF THE BID.

SA GERRY M. PERDIDO
Chairman, BAC Secretariat